

Clinical Performance Survey

Name: Colwood Back to Back Chiropractic Inc. - Dr. Julia Martorana

Overall outcomes* for :				
	Disc Herniation	Disc Degenerations	Sciatica Radiculopathy	Overall
86% or More				
71% to 85%				
56% to 70%				
41% to 55%				
Less than 40%				

Comments:

* The above data is not peer reviewed. It is the subjective response from a Physician grading their clinical results with the SpineMED® System.



Dear SpineMED® Operator:

As part of our ongoing effort to continually improve the SpineMED® System, we are collecting clinical performance data from all users of the device.

Please complete this survey and fax it to 250-563-3177 or email a copy to info@spinemed.com.

Thank you for your support.

Best regards,

Tim Emsky

CLINIC INFORMATION	
Physician Name: Dr. Julia Martorana	
Clinic Name: Colwood Back to Back Chiropractic Inc.	
Street Address: 591 Ledsham Rd.	
City: Victoria	
Province/State: BC	Postal Code/Zip: V9C1J9
Telephone: 250-391-8761	Email: back2back@shaw.ca
SpineMED® Serial #:	

CLINICAL PERFORMANCE SURVEY

- Please rate your overall patient outcomes for patients with Disc Herniations. (choose one)
 Less than 40% 41% to 55% 56% to 70% 71% to 85% 86% or more
- Please rate your overall patient outcomes for patients with Disc Degenerations. (choose one)
 Less than 40% 41% to 55% 56% to 70% 71% to 85% 86% or more
- Please rate your overall patient outcomes for patients with Sciatica Radiculopathy. (choose one)
 Less than 40% 41% to 55% 56% to 70% 71% to 85% 86% or more
- Please rate overall patient satisfaction with SpineMED® treatment. (choose one)
 Extremely Unsatisfied Unsatisfied Neutral Satisfied Extremely Satisfied

Additional Comments:



END USER USABILITY SURVEY

Dear SpineMED® Operator:

As part of our ongoing effort to continually improve the SpineMED® System, we are collecting usability data from all users of the device. The information we seek is how the SpineMED® System User Interface and Interaction performs in a real-world environment inclusive of human factors. We wish to gain your experience with any deficiencies or limitations you have encountered with regard to the usability of the device concerning staff operation of the device and also patients experience during treatment.

Please complete this survey and fax it to 250-563-3177 or email a copy to info@spinemed.com.

Thank you for your support.

Best regards,

Tim Emsky

CLINIC INFORMATION			
Physician Name:	Dr. Julia Martorana		
Clinic Name:	Back to Back Chiropractic Inc.		
Address:	591 ledsham Rd.		
City:	Victoria	Province/State:	BC
Postal Code/Zip:	V9C1J9	Email Address:	back2back@shaw.ca
SpineMED® Serial #:			

1. Please provide your observations or comments related to deficiencies in the usability of the device as a clinician operating the device.

Sometimes we have to get the patient off the table and restart the whole system and try again when it wont work. Also the on/off button on the table gets stuck alot

2. Please provide any suggestions for improving the usability of the device as a clinician operating the device.

Maybe change the on/off button to a switch or different type of button

3. Please provide your observations or comments related to deficiencies in the usability of the device from a patient's perspective.

4. Please provide any suggestions for improving the usability of the device from a patient's perspective.

Patients just complain about the table being too hard to lay on.